



LEED ID+C: Commercial Interiors | v3 - LEED 2009

Medical and process equipment efficiency

EApC3 | Possible 1 point

Glossary

Intent

To reduce energy consumption by using efficient medical and other equipment.

Requirements

Reduce the energy load for all diagnostic imaging equipment (x-rays, MRIs, etc), sterilization, and point-of-use electric steam generators installed in the project, (but excluding other types of medical equipment):

- At least 50%, by rated power, of medical equipment purchased for the project shall be among the 25th percentile of lowest energy consumers for that class of equipment. Equipment shall be compared based on their continuous (or "standby") mode electrical energy consumption.

Submittals

General

[REGISTER FOR THE PILOT CREDIT](#)

- Participate in the [LEEDuser pilot credit forum](#)
- Complete the feedback survey:

[CREDITS 1-14](#)

[CREDITS 15-27](#)

[CREDITS 28-42](#)

[CREDITS 43-56](#)

[CREDITS 57-67](#)

[CREDITS 68-82](#)

[CREDITS 83-103](#)